

# **Revolock**<sup>®</sup> V2 BOA<sup>®</sup> Dial and HC Lace Kit

Port A

RP-310A





WARRANTY: www.clickmedical.co/terms/#warranty U.S. Customers: Contact Click Medical directly Help@ClickMedical.co Help Line: +1 970 670 7012 International Customers:

Please contact your local distributor

Click Medical products are patent protected. For full list see www.clickmedical.co/patents

Revised 2023



### Advisory

Must be installed by a professional trained in RevoLock fabrication. Lace will wear. At minimum, inspect monthly and replace every 6 months. Loss of suspension can occur if lace fails. Do not use on patients with nerve issues. Keep lace away from open flame and sharp edges. Secure insert to liner before tensioning system.

See advisories in other languages: clickmedical.co/advisories

## **RevoLock**<sup>•</sup> V2 BOA Dial and HC Lace Kit Port A Contents:

Lace





BOA® Lanyard V2 Dial

Lace Feeder

For video education on Click Adjustability and RevoLock operation **Visit www.clickacademy.co** 

## Remove dial.

- a) Locate the metal insert.
- b) Insert a small flathead screwdriver between the metal insert and the dial body.
- c) Gently pry the reel upwards.



MDSS-UK RP LIMITED, 6 Wilmslow Road Rusholme, M14 5TP Manchester United Kingdom

CH REP

MDSS CH GmbH, Laurenzenvorstadt 61 5000 Aarau, Switzerland




MDSS GmbH, Schiffgraben 41 30175 Hannover, Germany






Click Medical, LLC, 1205 Hilltop Parkway, W101 Steamboat Springs, CO 80487, USA +1-970-670-7012




#### Remove the lace.

a) Ensure dial is open by pulling it up.



b) Pull all remaining lace out of the dial.



- c) Locate the large lace exit opening in the dial base.
  Hook lace loop in dial with a dental pick and pull lace out of dial.
- d) Cut lace and pull it out of the dial. Then remove lace from the device.

Confirm that you are using the correct replacement reel by identifying the C on the bottom of the dial.



